

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV013121\
 Data File : VV020226.D
 Acq On : 29 Jan 2021 16:04
 Operator : SY/MD
 Sample : VSTD00233
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD002210

Quant Time: Jan 30 03:39:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM013121.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jan 30 03:37:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)
Internal Standards					
1) 1,4-Difluorobenzene	5.639	114	13664	0.50 ug/L	0.00
5) Chlorobenzene-d5	8.863	117	13085	0.50 ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	6897	0.50 ug/L	0.00
System Monitoring Compounds					
2) Vinyl Chloride-d3	1.303	65	33129	2.65 ug/L	0.00
4) 1,2-Dichloroethane-d4	5.042	65	20773	1.88 ug/L	0.00
7) 1,2-Dichloropropane-d6	6.068	67	24358	1.87 ug/L	0.00
8) Toluene-d8	7.317	98	68593	2.20 ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.216	84	14644	1.91 ug/L	0.00
Target Compounds					
3) Vinyl chloride	1.310	62	30815	1.61 ug/L	Qvalue 95
6) Trichloroethene	5.925	95	21384	1.67 ug/L	99
9) 1,2-Dibromoethane	8.353	107	12541	1.83 ug/L #	93
12) 1,2,3-Trichloropropane	10.267	75	11686	1.65 ug/L	100
13) 1,2-Dibromo-3-chloropr...	12.426	75	1833	1.80 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV013121\
Data File : VV020226.D
Acq On : 29 Jan 2021 16:04
Operator : SY/MD
Sample : VSTD00233
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD002210

Quant Time: Jan 30 03:39:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM013121.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jan 30 03:37:50 2021
Response via : Initial Calibration

