

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV013121\
 Data File : VV020227.D
 Acq On : 29 Jan 2021 17:06
 Operator : SY/MD
 Sample : VSTDICV0.5
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VICV211

Quant Time: Jan 30 03:49:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM013121.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jan 30 03:45:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	14412	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.863	117	13626	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	7141	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.303	65	8336	0.46	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery =	92.00%		
4) 1,2-Dichloroethane-d4	5.042	65	5174	0.47	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	94.00%		
7) 1,2-Dichloropropane-d6	6.068	67	6090	0.49	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery =	98.00%		
8) Toluene-d8	7.317	98	16731	0.48	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	96.00%		
10) 1,1,2,2-Tetrachloroeth...	10.216	84	3458	0.46	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery =	92.00%		
Target Compounds						
						Qvalue
3) Vinyl chloride	1.310	62	7476	0.46	ug/L	94
6) Trichloroethene	5.925	95	5108	0.47	ug/L	100
9) 1,2-Dibromoethane	8.368	107	2934	0.46	ug/L	100
12) 1,2,3-Trichloropropane	10.267	75	2764	0.46	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.426	75	435	0.42	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV013121\
Data File : VV020227.D
Acq On : 29 Jan 2021 17:06
Operator : SY/MD
Sample : VSTDICV0.5
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VICV211

Quant Time: Jan 30 03:49:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM013121.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jan 30 03:45:01 2021
Response via : Initial Calibration

