

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW013123\
 Data File : VW030005.D
 Acq On : 31 Jan 2023 11:05
 Operator : SY/MD
 Sample : VW0131WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK300

Quant Time: Feb 01 00:14:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Feb 01 00:13:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.606	114	238124	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.841	117	223379	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.236	152	111696	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	84001	4.651	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.000%
7) Chloroethane-d5	1.565	69	78717	5.258	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.200%
11) 1,1-Dichloroethene-d2	2.102	63	115233	3.500	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.000%
20) 2-Butanone-d5	3.886	46	156618	48.924	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.840%
24) Chloroform-d	4.336	84	147382	4.448	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.000%
26) 1,2-Dichloroethane-d4	5.024	65	64835	4.621	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.400%
32) Benzene-d6	5.040	84	315167	4.756	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.200%
36) 1,2-Dichloropropane-d6	6.060	67	100156	5.165	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.200%
41) Toluene-d8	7.307	98	266291	4.341	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.800%
43) trans-1,3-Dichloroprop...	7.613	79	31655	4.469	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.400%
46) 2-Hexanone-d5	8.082	63	139358	48.332	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.660%
56) 1,1,2,2-Tetrachloroeth...	10.204	84	66042	4.647	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.000%
66) 1,2-Dichlorobenzene-d4	11.612	152	93766	4.617	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV013123\
Data File : VV030005.D
Acq On : 31 Jan 2023 11:05
Operator : SY/MD
Sample : VV0131WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK300

Quant Time: Feb 01 00:14:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Feb 01 00:13:14 2023
Response via : Initial Calibration

