

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW013124\
 Data File : VW033913.D
 Acq On : 31 Jan 2024 10:12
 Operator : SY/MD
 Sample : VW0131MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK227

Quant Time: Feb 01 05:28:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Feb 01 04:45:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	223805	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	208982	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	104171	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	70156	40.787	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.580%
7) Chloroethane-d5	1.532	69	58604	43.023	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.040%
11) 1,1-Dichloroethene-d2	2.060	65	32590	44.025	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.060%
21) 2-Butanone-d5	3.783	46	89605	75.950	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	75.950%
24) Chloroform-d	4.249	84	129092	41.972	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.940%
26) 1,2-Dichloroethane-d4	4.941	65	90246	45.413	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.820%
32) Benzene-d6	4.960	84	250384	42.144	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.280%
36) 1,2-Dichloropropane-d6	5.989	67	79775	40.792	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	81.580%
41) Toluene-d8	7.243	98	229174	42.812	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.620%
43) trans-1,3-Dichloroprop...	7.551	79	39803	41.181	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.360%
47) 2-Hexanone-d5	8.024	63	64491	79.501	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	79.500%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	103408	41.419	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	82.840%
66) 1,2-Dichlorobenzene-d4	11.561	152	89185	45.125	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.240%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV013124\
Data File : VV033913.D
Acq On : 31 Jan 2024 10:12
Operator : SY/MD
Sample : VV0131MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK227

Quant Time: Feb 01 05:28:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Feb 01 04:45:22 2024
Response via : Initial Calibration

