

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW013124\
 Data File : VW033921.D
 Acq On : 31 Jan 2024 15:29
 Operator : SY/MD
 Sample : P1303-06
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Feb 01 05:30:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Feb 01 04:45:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	176365	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	166155	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	81313	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	47087	34.739	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	69.480%
7) Chloroethane-d5	1.532	69	44211	41.187	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	82.380%
11) 1,1-Dichloroethene-d2	2.059	65	23003	39.433	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.860%
21) 2-Butanone-d5	3.783	46	90533	97.377	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.380%
24) Chloroform-d	4.249	84	109236	45.070	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.140%
26) 1,2-Dichloroethane-d4	4.940	65	83145	53.094	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.180%
32) Benzene-d6	4.960	84	205098	43.419	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.840%
36) 1,2-Dichloropropane-d6	5.989	67	68494	44.051	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	88.100%
41) Toluene-d8	7.243	98	178249	41.881	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.760%
43) trans-1,3-Dichloroprop...	7.551	79	34738	45.205	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.400%
47) 2-Hexanone-d5	8.021	63	60539	93.865	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	93.860%
56) 1,1,2,2-Tetrachloroeth...	10.152	84	96970	48.851	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.700%
66) 1,2-Dichlorobenzene-d4	11.561	152	71894	46.602	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV013124\
Data File : VV033921.D
Acq On : 31 Jan 2024 15:29
Operator : SY/MD
Sample : P1303-06
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Feb 01 05:30:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Feb 01 04:45:22 2024
Response via : Initial Calibration

