

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW013124\
 Data File : VW033928.D
 Acq On : 31 Jan 2024 18:16
 Operator : SY/MD
 Sample : P1309-07
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH907

Quant Time: Feb 01 05:32:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Feb 01 04:45:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	217748	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	202366	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	100822	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	61760	36.905	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	73.800%
7) Chloroethane-d5	1.529	69	51867	39.137	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	78.280%
11) 1,1-Dichloroethene-d2	2.060	65	29366	40.774	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.540%
21) 2-Butanone-d5	3.786	46	79589	69.337	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	69.340%
24) Chloroform-d	4.249	84	116595	38.964	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	77.920%
26) 1,2-Dichloroethane-d4	4.940	65	82033	42.428	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.860%
32) Benzene-d6	4.960	84	223587	38.864	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	77.720%
36) 1,2-Dichloropropane-d6	5.989	67	69596	36.751	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	73.500%
41) Toluene-d8	7.243	98	202206	39.009	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	78.020%#
43) trans-1,3-Dichloroprop...	7.554	79	35638	38.078	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.160%
47) 2-Hexanone-d5	8.024	63	58577	74.571	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	74.570%
56) 1,1,2,2-Tetrachloroeth...	10.152	84	94917	39.261	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	78.520%
66) 1,2-Dichlorobenzene-d4	11.561	152	80486	42.076	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.160%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV013124\
Data File : VV033928.D
Acq On : 31 Jan 2024 18:16
Operator : SY/MD
Sample : P1309-07
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH907

Quant Time: Feb 01 05:32:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Feb 01 04:45:22 2024
Response via : Initial Calibration

