

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW020124\
 Data File : VW033947.D
 Acq On : 01 Feb 2024 14:38
 Operator : SY/MD
 Sample : P1310-06
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH912

Quant Time: Feb 02 04:31:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Feb 02 04:28:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	204776	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	187966	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	94535	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	56181	35.698	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	71.400%
7) Chloroethane-d5	1.529	69	48355	38.798	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	77.600%
11) 1,1-Dichloroethene-d2	2.060	65	26770	39.524	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.040%
21) 2-Butanone-d5	3.783	46	88070	81.585	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	81.590%
24) Chloroform-d	4.249	84	113487	40.327	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.660%
26) 1,2-Dichloroethane-d4	4.941	65	81979	45.086	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.180%
32) Benzene-d6	4.960	84	216879	40.586	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.180%
36) 1,2-Dichloropropane-d6	5.989	67	68526	38.958	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	77.920%
41) Toluene-d8	7.243	98	195768	40.660	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	81.320%
43) trans-1,3-Dichloroprop...	7.555	79	35613	40.966	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.940%
47) 2-Hexanone-d5	8.024	63	58822	80.620	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	80.620%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	96548	42.995	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	85.980%
66) 1,2-Dichlorobenzene-d4	11.561	152	79764	44.472	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.940%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV020124\
Data File : VV033947.D
Acq On : 01 Feb 2024 14:38
Operator : SY/MD
Sample : P1310-06
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH912

Quant Time: Feb 02 04:31:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Feb 02 04:28:29 2024
Response via : Initial Calibration

