

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW020124\
 Data File : VW033949.D
 Acq On : 01 Feb 2024 15:26
 Operator : SY/MD
 Sample : P1310-08
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH920

Quant Time: Feb 02 04:32:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Feb 02 04:28:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	204507	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	191185	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	95543	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	54030	34.376	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	68.760%
7) Chloroethane-d5	1.529	69	47119	37.856	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	75.720%
11) 1,1-Dichloroethene-d2	2.060	65	26354	38.961	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.920%
21) 2-Butanone-d5	3.783	46	86283	80.035	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	80.040%
24) Chloroform-d	4.249	84	110682	39.382	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	78.760%
26) 1,2-Dichloroethane-d4	4.941	65	77987	42.947	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.900%
32) Benzene-d6	4.960	84	210647	38.756	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	77.520%
36) 1,2-Dichloropropane-d6	5.989	67	67343	37.641	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	75.280%
41) Toluene-d8	7.243	98	190688	38.938	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	77.880%#
43) trans-1,3-Dichloroprop...	7.555	79	34837	39.399	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.800%
47) 2-Hexanone-d5	8.024	63	56805	76.544	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	76.540%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	92737	40.602	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	81.200%
66) 1,2-Dichlorobenzene-d4	11.561	152	76191	42.031	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.060%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV020124\
Data File : VV033949.D
Acq On : 01 Feb 2024 15:26
Operator : SY/MD
Sample : P1310-08
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH920

Quant Time: Feb 02 04:32:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Feb 02 04:28:29 2024
Response via : Initial Calibration

