

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW020124\
 Data File : VW033953.D
 Acq On : 01 Feb 2024 17:02
 Operator : SY/MD
 Sample : P1300-09
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH6Y3

Quant Time: Feb 02 04:33:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Feb 02 04:28:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	179744	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	173255	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	84246	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	40380	29.231	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	58.460%#
7) Chloroethane-d5	1.526	69	39100	35.741	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	71.480%
11) 1,1-Dichloroethene-d2	2.056	65	20023	33.679	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.360%
21) 2-Butanone-d5	3.783	46	89018	93.948	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	93.950%
24) Chloroform-d	4.249	84	100490	40.682	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.360%
26) 1,2-Dichloroethane-d4	4.941	65	73506	46.056	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.120%
32) Benzene-d6	4.960	84	184693	37.497	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	75.000%
36) 1,2-Dichloropropane-d6	5.989	67	60846	37.529	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	75.060%
41) Toluene-d8	7.243	98	162260	36.562	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	73.120%#
43) trans-1,3-Dichloroprop...	7.555	79	33007	41.192	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.380%
47) 2-Hexanone-d5	8.021	63	60410	89.826	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	89.830%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	92209	44.549	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.100%
66) 1,2-Dichlorobenzene-d4	11.561	152	68497	42.854	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.700%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV020124\
Data File : VV033953.D
Acq On : 01 Feb 2024 17:02
Operator : SY/MD
Sample : P1300-09
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH6Y3

Quant Time: Feb 02 04:33:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Feb 02 04:28:29 2024
Response via : Initial Calibration

