

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW020124\
 Data File : VW033954.D
 Acq On : 01 Feb 2024 17:26
 Operator : SY/MD
 Sample : P1300-10
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH700

Quant Time: Feb 02 04:33:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Feb 02 04:28:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	207640	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	192502	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	94798	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	52086	32.639	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	65.280%	
7) Chloroethane-d5	1.529	69	45342	35.879	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	71.760%	
11) 1,1-Dichloroethene-d2	2.060	65	25579	37.244	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	74.480%	
21) 2-Butanone-d5	3.783	46	86294	78.838	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	78.840%	
24) Chloroform-d	4.249	84	110516	38.730	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	77.460%	
26) 1,2-Dichloroethane-d4	4.940	65	76381	41.428	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	82.860%	
32) Benzene-d6	4.960	84	201951	36.902	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	73.800%	
36) 1,2-Dichloropropane-d6	5.989	67	64909	36.032	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	72.060%	
41) Toluene-d8	7.243	98	180837	36.674	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	73.340%#	
43) trans-1,3-Dichloroprop...	7.554	79	33369	37.480	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	74.960%	
47) 2-Hexanone-d5	8.024	63	60138	80.481	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	80.480%	
56) 1,1,2,2-Tetrachloroeth...	10.152	84	92253	40.114	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	80.220%	
66) 1,2-Dichlorobenzene-d4	11.561	152	74485	41.413	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	82.820%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV020124\
Data File : VV033954.D
Acq On : 01 Feb 2024 17:26
Operator : SY/MD
Sample : P1300-10
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH700

Quant Time: Feb 02 04:33:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Feb 02 04:28:29 2024
Response via : Initial Calibration

