

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW020223\
 Data File : VW030023.D
 Acq On : 02 Feb 2023 15:08
 Operator : SY/MD
 Sample : 01408-01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C1156

Quant Time: Feb 03 01:14:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Feb 03 01:11:57 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	229496	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	213989	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.236	152	107959	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	79058	4.542	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.800%
7) Chloroethane-d5	1.564	69	76802	5.323	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	106.400%
11) 1,1-Dichloroethene-d2	2.101	63	113270	3.569	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.400%
20) 2-Butanone-d5	3.886	46	164573	53.342	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.680%
24) Chloroform-d	4.336	84	150928	4.727	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.600%
26) 1,2-Dichloroethane-d4	5.024	65	65403	4.836	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.800%
32) Benzene-d6	5.040	84	318281	5.014	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.200%
36) 1,2-Dichloropropane-d6	6.059	67	101799	5.480	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	109.600%
41) Toluene-d8	7.307	98	267373	4.550	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.000%
43) trans-1,3-Dichloroprop...	7.612	79	31943	4.708	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.200%
46) 2-Hexanone-d5	8.082	63	149190	54.012	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	108.020%
56) 1,1,2,2-Tetrachloroeth...	10.204	84	69552	5.109	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.200%
66) 1,2-Dichlorobenzene-d4	11.612	152	98194	5.003	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.000%
Target Compounds						Qvalue
13) Acetone	2.185	43	6576	2.986	ug/L	68
16) Methylene chloride	2.500	84	5771	0.237	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV020223\
Data File : VV030023.D
Acq On : 02 Feb 2023 15:08
Operator : SY/MD
Sample : 01408-01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C1156

Quant Time: Feb 03 01:14:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Feb 03 01:11:57 2023
Response via : Initial Calibration

