

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW020224\
 Data File : VW033959.D
 Acq On : 02 Feb 2024 10:07
 Operator : SY/MD
 Sample : VW0202WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK231

Quant Time: Feb 03 03:34:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Feb 03 01:42:15 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	204803	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	189198	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	95439	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	71121	45.185	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.360%
7) Chloroethane-d5	1.529	69	56870	45.624	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	91.240%
11) 1,1-Dichloroethene-d2	2.060	65	32565	48.073	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.140%
21) 2-Butanone-d5	3.783	46	84993	78.725	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	78.720%
24) Chloroform-d	4.249	84	117054	41.589	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.180%
26) 1,2-Dichloroethane-d4	4.941	65	83009	45.646	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.300%
32) Benzene-d6	4.960	84	231761	43.088	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.180%
36) 1,2-Dichloropropane-d6	5.989	67	69637	39.332	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	78.660%
41) Toluene-d8	7.243	98	214492	44.259	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.520%
43) trans-1,3-Dichloroprop...	7.555	79	36154	41.318	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.640%
47) 2-Hexanone-d5	8.024	63	55848	76.045	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	76.050%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	91352	40.416	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	80.840%
66) 1,2-Dichlorobenzene-d4	11.561	152	80647	44.538	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.080%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV020224\
Data File : VV033959.D
Acq On : 02 Feb 2024 10:07
Operator : SY/MD
Sample : VV0202WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK231

Quant Time: Feb 03 03:34:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Feb 03 01:42:15 2024
Response via : Initial Calibration

