

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW020224\
 Data File : VW033962.D
 Acq On : 02 Feb 2024 11:18
 Operator : SY/MD
 Sample : P1300-11
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Feb 03 03:34:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Feb 03 01:42:15 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	198918	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	185611	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	92887	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	72961	47.725	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.460%
7) Chloroethane-d5	1.529	69	57137	47.194	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.380%
11) 1,1-Dichloroethene-d2	2.060	65	32426	49.284	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.560%
21) 2-Butanone-d5	3.783	46	83330	79.468	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	79.470%
24) Chloroform-d	4.249	84	118401	43.313	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.620%
26) 1,2-Dichloroethane-d4	4.941	65	81008	45.864	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.720%
32) Benzene-d6	4.960	84	232333	44.029	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.060%
36) 1,2-Dichloropropane-d6	5.989	67	70391	40.526	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	81.060%
41) Toluene-d8	7.243	98	213028	44.806	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.620%
43) trans-1,3-Dichloroprop...	7.555	79	36178	42.144	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.280%
47) 2-Hexanone-d5	8.024	63	56329	78.182	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	78.180%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	90934	41.009	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	82.020%
66) 1,2-Dichlorobenzene-d4	11.561	152	80247	45.535	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.060%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV020224\
Data File : VV033962.D
Acq On : 02 Feb 2024 11:18
Operator : SY/MD
Sample : P1300-11
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Feb 03 03:34:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Feb 03 01:42:15 2024
Response via : Initial Calibration

