

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW020224\
 Data File : VW033963.D
 Acq On : 02 Feb 2024 11:42
 Operator : SY/MD
 Sample : P1329-17
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Feb 03 03:35:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Feb 03 01:42:15 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	193887	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	181384	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	89278	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	71480	47.970	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.940%
7) Chloroethane-d5	1.529	69	55698	47.199	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.400%
11) 1,1-Dichloroethene-d2	2.060	65	31698	49.428	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.860%
21) 2-Butanone-d5	3.783	46	79191	77.480	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	77.480%
24) Chloroform-d	4.249	84	115819	43.467	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.940%
26) 1,2-Dichloroethane-d4	4.941	65	81140	47.131	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.260%
32) Benzene-d6	4.960	84	226831	43.988	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.980%
36) 1,2-Dichloropropane-d6	5.989	67	68545	40.383	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	80.760%
41) Toluene-d8	7.243	98	207663	44.696	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.400%
43) trans-1,3-Dichloroprop...	7.555	79	36081	43.010	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.020%
47) 2-Hexanone-d5	8.024	63	54521	77.436	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	77.440%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	91102	42.042	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	84.080%
66) 1,2-Dichlorobenzene-d4	11.561	152	79479	46.922	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.840%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV020224\
Data File : VV033963.D
Acq On : 02 Feb 2024 11:42
Operator : SY/MD
Sample : P1329-17
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Feb 03 03:35:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Feb 03 01:42:15 2024
Response via : Initial Calibration

