

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW020224\
 Data File : VW033964.D
 Acq On : 02 Feb 2024 12:12
 Operator : SY/MD
 Sample : P1309-16
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Feb 03 03:35:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Feb 03 01:42:15 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	183568	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	173020	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	88136	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	65519	46.441	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	92.880%
7) Chloroethane-d5	1.529	69	51522	46.115	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.220%
11) 1,1-Dichloroethene-d2	2.056	65	30304	49.910	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	99.820%
21) 2-Butanone-d5	3.783	46	77594	80.185	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	80.190%
24) Chloroform-d	4.246	84	109174	43.277	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.560%
26) 1,2-Dichloroethane-d4	4.941	65	79116	48.538	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.080%
32) Benzene-d6	4.957	84	216838	44.083	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.160%
36) 1,2-Dichloropropane-d6	5.989	67	65420	40.405	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	80.800%
41) Toluene-d8	7.243	98	200403	45.218	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.440%
43) trans-1,3-Dichloroprop...	7.555	79	33109	41.376	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.760%
47) 2-Hexanone-d5	8.024	63	52500	78.171	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	78.170%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	87670	42.414	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	84.820%
66) 1,2-Dichlorobenzene-d4	11.561	152	77815	46.535	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.060%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV020224\
Data File : VV033964.D
Acq On : 02 Feb 2024 12:12
Operator : SY/MD
Sample : P1309-16
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Feb 03 03:35:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Feb 03 01:42:15 2024
Response via : Initial Calibration

