

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV020320\
 Data File : VV014489.D
 Acq On : 03 Feb 2020 18:40
 Operator : SY/MD
 Sample : VIBLK26
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK26

Quant Time: Feb 04 06:05:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM020120WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Feb 04 05:31:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	445441	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	408818	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	195990	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	102882	41.59	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.18%
7) Chloroethane-d5	1.59	69	75945	49.37	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.74%
11) 1,1-Dichloroethene-d2	2.13	63	117247	35.50	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.00%
21) 2-Butanone-d5	3.92	46	167547	100.66	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	100.66%
24) Chloroform-d	4.40	84	238371	45.53	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.06%
26) 1,2-Dichloroethane-d4	5.07	65	152426	48.33	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.66%
32) Benzene-d6	5.09	84	511532	47.52	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.04%
36) 1,2-Dichloropropane-d6	6.11	67	162179	47.62	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.24%
41) Toluene-d8	7.35	98	473541	47.65	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.30%
43) trans-1,3-Dichloropropene-	7.66	79	83239	48.09	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.18%
47) 2-Hexanone-d5	8.13	63	131098	98.39	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	98.39%
57) 1,1,2,2-Tetrachloroethane-	10.25	84	203003	46.26	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.52%
64) 1,2-Dichlorobenzene-d4	11.67	152	175011	48.48	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.96%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV020320\
Data File : VV014489.D
Acq On : 03 Feb 2020 18:40
Operator : SY/MD
Sample : VIBLK26
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK26

Quant Time: Feb 04 06:05:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM020120WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Feb 04 05:31:51 2020
Response via : Initial Calibration

