

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW020321\
 Data File : VW020264.D
 Acq On : 03 Feb 2021 20:16
 Operator : SY/MD
 Sample : M1258-05
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Feb 03 23:53:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM020321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Feb 03 17:26:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.621	114	566312	50.00	ug/L	# 0.00
28) Chlorobenzene-d5	8.855	117	558522	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.254	152	308758	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.309	65	106976	47.12	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.24%
7) Chloroethane-d5	1.569	69	92720	47.98	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.96%
11) 1,1-Dichloroethene-d2	2.106	63	163813	35.38	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.76%
21) 2-Butanone-d5	3.894	46	181698	97.34	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.34%
24) Chloroform-d	4.354	84	287065	47.93	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.86%
26) 1,2-Dichloroethane-d4	5.036	65	181967	49.97	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.94%
32) Benzene-d6	5.055	84	581416	49.38	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.76%
36) 1,2-Dichloropropane-d6	6.074	67	171330	49.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	98.52%
41) Toluene-d8	7.318	98	562060	48.32	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.64%
43) trans-1,3-Dichloroprop...	7.624	79	88330	47.59	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.18%
47) 2-Hexanone-d5	8.090	63	146201	94.85	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	94.85%
57) 1,1,2,2-Tetrachloroeth...	10.222	84	262784	48.31	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.62%
64) 1,2-Dichlorobenzene-d4	11.630	152	288086	50.54	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.08%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV020321\
Data File : VV020264.D
Acq On : 03 Feb 2021 20:16
Operator : SY/MD
Sample : M1258-05
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Feb 03 23:53:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM020321WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Feb 03 17:26:57 2021
Response via : Initial Calibration

