

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW020323\
 Data File : VW030035.D
 Acq On : 03 Feb 2023 12:11
 Operator : SY/MD
 Sample : 01408-05
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Feb 04 01:03:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Feb 04 01:01:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.606	114	229641	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.841	117	225432	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.236	152	112085	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	79252	4.550	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.000%
7) Chloroethane-d5	1.565	69	77333	5.356	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.200%
11) 1,1-Dichloroethene-d2	2.101	63	111941	3.525	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.600%
20) 2-Butanone-d5	3.896	46	171316	55.492	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.980%
24) Chloroform-d	4.336	84	149195	4.669	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.400%
26) 1,2-Dichloroethane-d4	5.021	65	66686	4.928	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.600%
32) Benzene-d6	5.040	84	315633	4.720	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.400%
36) 1,2-Dichloropropane-d6	6.060	67	101102	5.166	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.400%
41) Toluene-d8	7.307	98	258240	4.171	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.400%
43) trans-1,3-Dichloroprop...	7.612	79	30743	4.301	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.000%
46) 2-Hexanone-d5	8.082	63	152577	52.435	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.860%
56) 1,1,2,2-Tetrachloroeth...	10.204	84	71721	5.001	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.000%
66) 1,2-Dichlorobenzene-d4	11.609	152	98400	4.829	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV020323\
Data File : VV030035.D
Acq On : 03 Feb 2023 12:11
Operator : SY/MD
Sample : 01408-05
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Feb 04 01:03:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Feb 04 01:01:21 2023
Response via : Initial Calibration

