

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW020624\
 Data File : VW034011.D
 Acq On : 06 Feb 2024 15:31
 Operator : SY/MD
 Sample : P1289-12DL 100X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AF5DL

Quant Time: Feb 07 04:41:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Feb 07 00:49:42 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	93730	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	86727	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	42814	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	24879	3.427	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	68.600%
7) Chloroethane-d5	1.532	69	23736	3.733	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	74.600%
11) 1,1-Dichloroethene-d2	2.060	65	13295	3.798	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.000%
20) 2-Butanone-d5	3.834	46	44503	36.503	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	73.000%
24) Chloroform-d	4.252	84	52012	3.943	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.800%
26) 1,2-Dichloroethane-d4	4.947	65	25277	4.100	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.000%
32) Benzene-d6	4.963	84	99403	3.820	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	76.400%
36) 1,2-Dichloropropane-d6	5.992	67	28877	3.970	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	79.400%
41) Toluene-d8	7.246	98	93021	3.890	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.800%
43) trans-1,3-Dichloroprop...	7.561	79	10209	3.717	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.400%
46) 2-Hexanone-d5	8.030	63	41461	42.289	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.580%
56) 1,1,2,2-Tetrachloroeth...	10.152	84	18328	4.156	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.200%
66) 1,2-Dichlorobenzene-d4	11.561	152	32190	4.134	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	82.600%
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.069	96	1888	0.288	ug/L	# 1
19) 1,1-Dichloroethane	3.117	63	9056	0.739	ug/L	97
29) 1,1,1-Trichloroethane	4.513	97	43226	3.712	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV020624\
Data File : VV034011.D
Acq On : 06 Feb 2024 15:31
Operator : SY/MD
Sample : P1289-12DL 100X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AF5DL

Quant Time: Feb 07 04:41:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Feb 07 00:49:42 2024
Response via : Initial Calibration

