



Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV020724\
Data File : VV034017.D
Acq On : 07 Feb 2024 08:27
Operator : SY/MD
Sample : BFB247
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFB247

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Title : TRACE VOA SFAM1.0
Last Update : Fri Feb 09 03:48:51 2024



AutoFind: Scans 2789, 2790, 2791; Background Corrected with Scan 2777

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.9	12227	PASS
75	95	30	80	58.0	32349	PASS
95	95	100	100	100.0	55808	PASS
96	95	5	9	6.3	3515	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	85.4	47653	PASS
175	174	5	9	7.9	3787	PASS
176	174	95	105	97.9	46664	PASS
177	176	5	10	6.6	3087	PASS

m/z	Abundance
43.90	140.0

Instrument :
MSVOA_V
ClientSampleId :
BFB247

m/z	Abundance
35.80	660.0
37.00	3188.0
37.90	2800.0
38.90	1019.0
39.80	144.0
43.90	495.0
44.90	589.0
46.80	772.0
47.90	552.0
48.90	2800.0
49.90	12660.0
50.90	3687.0
51.90	145.0
54.90	171.0
55.90	1216.0
56.90	1940.0
59.80	632.0
60.90	3229.0
61.90	3465.0
62.90	2639.0
64.00	282.0
64.80	131.0
66.80	247.0
67.90	6404.0
68.90	6896.0
69.90	530.0
71.90	393.0
72.90	3177.0
74.00	11194.0
74.90	33104.0
75.90	3088.0
76.90	293.0
77.60	201.0
78.90	2620.0
79.80	763.0
80.90	2869.0
81.90	681.0
86.90	1984.0
87.90	2168.0
90.80	289.0
91.90	2010.0
92.90	2668.0
94.00	7726.0
95.00	56408.0
95.90	3804.0
103.70	273.0
104.90	104.0
105.80	349.0
110.70	133.0
112.80	115.0
115.80	319.0
116.80	453.0
117.90	294.0
118.80	422.0
127.80	250.0
128.80	116.0
129.80	224.0
130.90	109.0
134.80	146.0
136.80	128.0
140.80	929.0
141.70	161.0
142.80	893.0
145.70	118.0
147.90	192.0
154.80	139.0
156.80	126.0
160.80	105.0
171.90	255.0
173.90	46976.0
175.00	3789.0
175.90	46072.0
176.90	3100.0

Instrument :
MSVOA_V
ClientSampleId :
BFB247

m/z	Abundance
35.80	715.0
37.00	3163.0
37.90	2763.0
38.90	977.0
39.80	136.0
43.90	506.0
44.90	609.0
46.80	713.0
47.90	521.0
48.90	2867.0
49.90	12673.0
51.00	3670.0
51.90	140.0
54.90	168.0
55.90	1201.0
56.90	1886.0
59.80	598.0
60.90	3266.0
61.90	3377.0
62.90	2697.0
64.00	278.0
64.80	135.0
66.80	260.0
67.90	6468.0
68.90	6929.0
69.90	527.0
71.80	394.0
72.90	3133.0
73.90	11435.0
74.90	33512.0
75.90	3258.0
77.00	291.0
77.70	235.0
78.90	2609.0
79.80	788.0
80.90	2915.0
81.90	706.0
86.90	2082.0
87.90	2302.0
90.80	301.0
92.00	2001.0
92.90	2654.0
93.90	7874.0
95.00	57816.0
95.90	3653.0
103.70	312.0
104.80	111.0
105.80	353.0
110.70	149.0
112.70	108.0
115.80	310.0
116.80	467.0
117.80	287.0
118.80	455.0
127.80	262.0
128.70	108.0
129.90	171.0
130.90	109.0
134.80	161.0
136.80	152.0
140.80	906.0
141.70	141.0
142.80	925.0
145.70	116.0
147.90	194.0
154.80	156.0
156.80	134.0
173.90	49432.0
174.90	3914.0
175.90	48480.0
176.90	3231.0

Instrument :
MSVOA_V
ClientSampleId :
BFB247

m/z	Abundance
35.80	596.0
37.00	2870.0
37.90	2453.0
38.90	869.0
39.80	116.0
43.90	478.0
44.90	541.0
46.80	578.0
47.90	413.0
48.90	2487.0
49.90	11349.0
50.90	3276.0
51.80	126.0
54.80	171.0
55.90	1097.0
56.90	1734.0
59.90	526.0
60.90	3069.0
61.90	2985.0
63.00	2447.0
63.90	243.0
64.90	111.0
66.80	232.0
67.90	5873.0
68.90	6306.0
69.90	458.0
71.80	358.0
72.90	2807.0
73.90	10721.0
75.00	30432.0
76.00	3048.0
76.90	270.0
77.70	256.0
78.90	2389.0
79.80	739.0
80.90	2685.0
81.90	623.0
86.90	1964.0
87.90	2309.0
90.80	285.0
92.00	1829.0
92.90	2461.0
93.90	7322.0
95.00	53200.0
95.90	3089.0
103.70	320.0
105.80	333.0
110.70	118.0
115.80	284.0
116.80	439.0
117.80	246.0
118.80	428.0
127.80	237.0
128.70	107.0
129.90	159.0
134.70	135.0
136.80	152.0
140.80	810.0
141.70	107.0
142.80	813.0
145.70	107.0
147.90	188.0
154.80	149.0
171.90	243.0
173.90	46552.0
174.90	3659.0
175.90	45440.0
176.90	2931.0

Instrument :
MSVOA_V
ClientSampleId :
BFB247