

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV020820\
 Data File : VV014532.D
 Acq On : 07 Feb 2020 12:33
 Operator : SY/MD
 Sample : VV0208MBL01
 Misc : 5.0µ/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK204

Quant Time: Feb 08 00:36:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVLM020620W.M
 Quant Title : VOC Analysis
 QLast Update : Sat Feb 08 00:03:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	340505	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	322002	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.29	152	160266	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	81449	44.11	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.22%
7) Chloroethane-d5	1.58	69	58424	45.91	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	91.82%
11) 1,1-Dichloroethene-d2	2.13	63	99145	35.36	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.72%
21) 2-Butanone-d5	3.92	46	153546	107.72	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	107.72%
24) Chloroform-d	4.40	84	205077	48.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.48%
26) 1,2-Dichloroethane-d4	5.08	65	134931	52.53	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.06%
32) Benzene-d6	5.10	84	443319	49.72	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.44%
36) 1,2-Dichloropropane-d6	6.11	67	138359	50.61	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.22%
41) Toluene-d8	7.36	98	416876	50.78	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.56%
43) trans-1,3-Dichloropropene-	7.66	79	72564	49.50	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	99.00%
47) 2-Hexanone-d5	8.13	63	128149	110.17	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	110.17%
56) 1,1,2,2-Tetrachloroethane-	10.25	84	185501	52.25	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	104.50%
66) 1,2-Dichlorobenzene-d4	11.67	152	159530	51.90	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV020820\
Data File : VV014532.D
Acq On : 07 Feb 2020 12:33
Operator : SY/MD
Sample : VV0208MBL01
Misc : 5.0µL/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK204

Quant Time: Feb 08 00:36:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVLM020620W.M
Quant Title : VOC Analysis
QLast Update : Sat Feb 08 00:03:07 2020
Response via : Initial Calibration

