

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV020821\
 Data File : VV020279.D
 Acq On : 08 Feb 2021 10:59
 Operator : SY/MD
 Sample : VV0208MBL01
 Misc : 5.00G/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK93

Quant Time: Feb 09 01:47:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM020521WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Feb 09 01:47:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	620348	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	600442	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.255	152	294001	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	200974	44.43	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.86%
7) Chloroethane-d5	1.568	69	173473	46.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.50%
11) 1,1-Dichloroethene-d2	2.108	63	295890	34.61	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.22%
21) 2-Butanone-d5	3.886	46	246912	97.72	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.72%
24) Chloroform-d	4.352	84	355744	44.20	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.40%
26) 1,2-Dichloroethane-d4	5.037	65	238471	46.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.48%
32) Benzene-d6	5.053	84	726345	46.69	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.38%
36) 1,2-Dichloropropane-d6	6.072	67	222587	47.93	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.86%
41) Toluene-d8	7.320	98	661715	45.26	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.52%
43) trans-1,3-Dichloroprop...	7.625	79	102332	43.23	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.46%
47) 2-Hexanone-d5	8.091	63	174745	90.44	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.44%
57) 1,1,2,2-Tetrachloroeth...	10.223	84	299524	44.91	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.82%
64) 1,2-Dichlorobenzene-d4	11.631	152	270482	48.62	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.24%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV020821\
Data File : VV020279.D
Acq On : 08 Feb 2021 10:59
Operator : SY/MD
Sample : VV0208MBL01
Misc : 5.00G/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK93

Quant Time: Feb 09 01:47:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM020521WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Feb 09 01:47:10 2021
Response via : Initial Calibration

