

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW020924\
 Data File : VW034110.D
 Acq On : 09 Feb 2024 14:26
 Operator : SY/MD
 Sample : P1289-08
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Feb 09 23:14:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Feb 09 22:58:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	77797	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	71884	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	37515	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	26624	4.419	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.400%
7) Chloroethane-d5	1.532	69	22520	4.267	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	85.400%
11) 1,1-Dichloroethene-d2	2.060	65	13395	4.610	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	92.200%
20) 2-Butanone-d5	3.802	46	43305	42.795	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	85.600%
24) Chloroform-d	4.256	84	48314	4.413	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.200%
26) 1,2-Dichloroethane-d4	4.947	65	24561	4.800	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.000%
32) Benzene-d6	4.963	84	93988	4.358	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.200%
36) 1,2-Dichloropropane-d6	5.992	67	25199	4.180	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	83.600%
41) Toluene-d8	7.246	98	89617	4.521	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.400%
43) trans-1,3-Dichloroprop...	7.561	79	10366	4.554	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.000%
46) 2-Hexanone-d5	8.027	63	38179	46.982	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.960%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	17892	4.895	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.000%
66) 1,2-Dichlorobenzene-d4	11.561	152	31291	4.586	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV020924\
Data File : VV034110.D
Acq On : 09 Feb 2024 14:26
Operator : SY/MD
Sample : P1289-08
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Feb 09 23:14:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Feb 09 22:58:56 2024
Response via : Initial Calibration

