



Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV020924\
Data File : VV034096.D
Acq On : 09 Feb 2024 08:21
Operator : SY/MD
Sample : BFB249
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFB249

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Title : TRACE VOA SFAM1.0
Last Update : Fri Feb 09 03:48:51 2024



AutoFind: Scans 2789, 2790, 2791; Background Corrected with Scan 2778

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.9	11729	PASS
75	95	30	80	59.6	31896	PASS
95	95	100	100	100.0	53517	PASS
96	95	5	9	7.0	3730	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	86.0	46032	PASS
175	174	5	9	7.7	3561	PASS
176	174	95	105	98.6	45381	PASS
177	176	5	10	6.9	3113	PASS

m/z	Abundance
39.90	120.0
43.80	122.0

Instrument :
MSVOA_V
ClientSampleId :
BFB249

m/z	Abundance
35.90	473.0
36.90	2957.0
37.90	2621.0
38.90	1049.0
39.90	154.0
43.90	506.0
44.90	586.0
46.90	607.0
47.80	428.0
48.90	2528.0
50.00	11695.0
50.90	3595.0
51.80	147.0
54.80	189.0
55.90	980.0
56.90	1872.0
60.00	710.0
61.00	3237.0
61.90	3284.0
63.00	2218.0
63.90	191.0
65.00	125.0
66.90	225.0
67.90	6332.0
68.90	6578.0
69.90	564.0
71.90	481.0
72.90	2819.0
74.00	11506.0
75.00	32104.0
75.90	2620.0
76.90	277.0
77.80	199.0
78.90	2722.0
79.80	756.0
80.80	2896.0
81.80	568.0
83.00	103.0
86.80	1860.0
87.80	1819.0
91.00	246.0
91.90	1879.0
92.90	2707.0
94.00	7360.0
94.90	52952.0
95.90	3699.0
103.80	481.0
105.90	394.0
115.80	318.0
116.80	473.0
117.80	274.0
118.90	377.0
127.80	290.0
128.90	167.0
129.80	278.0
140.80	678.0
141.80	128.0
142.90	821.0
147.90	153.0
154.80	151.0
172.10	264.0
173.90	44784.0
174.90	3543.0
175.90	43560.0
176.90	2975.0

Instrument :
MSVOA_V
ClientSampleId :
BFB249

m/z	Abundance
35.90	548.0
36.90	2914.0
37.90	2710.0
38.90	1050.0
39.90	125.0
43.90	561.0
44.90	565.0
46.80	606.0
47.80	409.0
48.90	2599.0
50.00	12130.0
50.90	3551.0
51.90	138.0
54.80	211.0
55.90	1045.0
56.90	1913.0
59.90	670.0
60.90	3322.0
61.90	3390.0
63.00	2383.0
63.90	249.0
64.90	104.0
67.00	229.0
67.90	6467.0
68.90	6712.0
69.90	649.0
71.90	498.0
72.90	3047.0
73.90	11952.0
75.00	32952.0
75.90	2632.0
77.00	310.0
77.80	216.0
78.90	2785.0
79.90	806.0
80.80	2929.0
81.80	627.0
86.80	2095.0
87.80	1897.0
91.00	291.0
92.00	1985.0
92.90	2890.0
94.00	7713.0
94.90	55472.0
95.90	3871.0
103.80	488.0
105.90	383.0
110.70	103.0
115.80	361.0
116.80	433.0
117.80	291.0
118.80	430.0
127.70	317.0
128.80	140.0
129.80	285.0
130.90	112.0
134.70	108.0
136.80	113.0
140.80	740.0
141.70	109.0
142.90	949.0
147.80	219.0
154.80	221.0
156.90	122.0
160.80	117.0
171.80	273.0
173.90	47744.0
174.90	3666.0
175.90	47192.0
176.90	3249.0

Instrument :
MSVOA_V
ClientSampleId :
BFB249

m/z	Abundance
35.90	547.0
36.90	2597.0
37.90	2487.0
38.90	941.0
40.00	104.0
43.90	514.0
44.90	498.0
46.80	547.0
47.70	356.0
48.90	2377.0
50.00	11362.0
50.90	3134.0
52.00	133.0
54.80	210.0
55.90	1006.0
56.90	1759.0
59.90	619.0
60.90	3163.0
61.90	3165.0
63.00	2186.0
63.90	271.0
67.00	216.0
67.90	6054.0
69.00	6253.0
69.90	590.0
71.90	433.0
72.90	2976.0
73.90	11189.0
75.00	30632.0
75.90	2402.0
77.00	293.0
77.80	183.0
78.90	2593.0
79.90	784.0
80.80	2642.0
81.80	590.0
86.80	2040.0
87.80	1744.0
91.00	300.0
92.00	1882.0
92.90	2782.0
94.00	7465.0
94.90	52128.0
95.90	3621.0
103.80	401.0
105.80	351.0
107.00	101.0
112.70	100.0
115.80	357.0
116.90	399.0
117.80	274.0
118.80	431.0
127.70	309.0
128.80	111.0
129.80	221.0
130.90	118.0
134.70	109.0
136.80	118.0
140.80	793.0
142.90	988.0
147.80	238.0
154.80	209.0
156.80	134.0
160.80	108.0
171.80	259.0
173.90	45568.0
174.90	3475.0
175.90	45392.0
176.90	3116.0

Instrument :
MSVOA_V
ClientSampleId :
BFB249