

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW021022\
 Data File : VW024574.D
 Acq On : 10 Feb 2022 17:39
 Operator : SY/MD
 Sample : N1476-04
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFXZ3

Quant Time: Feb 11 23:56:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR020922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Feb 11 23:54:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	141664	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	139543	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	71373	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	49682	4.184	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.600%
7) Chloroethane-d5	1.568	69	44792	4.462	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.200%
11) 1,1-Dichloroethene-d2	2.108	63	80305	3.092	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.800%
20) 2-Butanone-d5	3.908	46	86845	44.055	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	88.100%
24) Chloroform-d	4.349	84	96704	4.627	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.600%
26) 1,2-Dichloroethane-d4	5.034	65	45320	4.779	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.600%
32) Benzene-d6	5.050	84	190965	4.823	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.400%
36) 1,2-Dichloropropane-d6	6.069	67	58666	4.764	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.200%
41) Toluene-d8	7.313	98	177151	4.797	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.000%
43) trans-1,3-Dichloroprop...	7.625	79	22809	4.587	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.800%
46) 2-Hexanone-d5	8.088	63	86717	49.495	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.000%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	34065	4.646	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.000%
66) 1,2-Dichlorobenzene-d4	11.622	152	53793	4.676	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.600%
Target Compounds						
					Qvalue	
25) Chloroform	4.381	83	10352	0.397	ug/L	91
47) Tetrachloroethene	7.973	164	129388	13.005	ug/L	91
53) m,p-xylene	9.146	106	1108	0.044	ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021022\
Data File : VV024574.D
Acq On : 10 Feb 2022 17:39
Operator : SY/MD
Sample : N1476-04
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFXZ3

Quant Time: Feb 11 23:56:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR020922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Feb 11 23:54:44 2022
Response via : Initial Calibration

