

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW021022\
 Data File : VW024575.D
 Acq On : 10 Feb 2022 18:02
 Operator : SY/MD
 Sample : N1476-05
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFXZ4

Quant Time: Feb 11 23:56:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR020922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Feb 11 23:54:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	130768	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	128798	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	65444	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	47260	4.311	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	86.200%	
7) Chloroethane-d5	1.568	69	43235	4.666	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	93.400%	
11) 1,1-Dichloroethene-d2	2.111	63	74806	3.120	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	62.400%	
20) 2-Butanone-d5	3.902	46	98244	53.990	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	107.980%	
24) Chloroform-d	4.349	84	90841	4.709	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	94.200%	
26) 1,2-Dichloroethane-d4	5.034	65	43495	4.969	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	99.400%	
32) Benzene-d6	5.050	84	185617	5.079	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	101.600%	
36) 1,2-Dichloropropane-d6	6.069	67	56299	4.953	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	99.000%	
41) Toluene-d8	7.317	98	166464	4.884	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	97.600%	
43) trans-1,3-Dichloroprop...	7.622	79	21760	4.741	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	94.800%	
46) 2-Hexanone-d5	8.088	63	87353	54.018	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	108.040%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	34146	5.046	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	101.000%	
66) 1,2-Dichlorobenzene-d4	11.622	152	52984	5.022	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	100.400%	
Target Compounds						
					Qvalue	
13) Acetone	2.198	43	13079	9.108	ug/L	82
19) 1,1-Dichloroethane	3.201	63	2574	0.112	ug/L	92
25) Chloroform	4.378	83	25716	1.070	ug/L	95
34) Trichloroethene	5.918	95	3232	0.234	ug/L	84
47) Tetrachloroethene	7.976	164	12307	1.340	ug/L	87
72) 1,2,3-Trichlorobenzene	13.744	180	1345	0.115	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021022\
Data File : VV024575.D
Acq On : 10 Feb 2022 18:02
Operator : SY/MD
Sample : N1476-05
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFXZ4

Quant Time: Feb 11 23:56:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR020922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Feb 11 23:54:44 2022
Response via : Initial Calibration

