

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VW021025\
 Data File : VW038665.D
 Acq On : 10 Feb 2025 14:00
 Operator : SY/MD
 Sample : Q1335-02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E29C4

Quant Time: Feb 14 11:39:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR020625WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Feb 14 11:36:33 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	145762	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	149456	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	61582	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	65854	4.37	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.40%
7) Chloroethane-d5	1.532	69	47489	4.18	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	83.60%
11) 1,1-Dichloroethene-d2	2.056	65	25955	4.12	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	82.40%
20) 2-Butanone-d5	3.780	46	161723	61.79	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	123.58%
24) Chloroform-d	4.246	84	103403	4.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.60%
26) 1,2-Dichloroethane-d4	4.937	65	54568	5.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.80%
32) Benzene-d6	4.953	84	202828	4.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.80%
36) 1,2-Dichloropropane-d6	5.982	67	63605	4.91	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.20%
41) Toluene-d8	7.239	98	164775	4.22	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.40%
43) trans-1,3-Dichloroprop...	7.561	79	18940	3.98	ug/L	0.01
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.60%
46) 2-Hexanone-d5	8.018	63	123782	67.19	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	134.38%#
56) 1,1,2,2-Tetrachloroeth...	10.146	84	61778	5.61	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	112.20%
66) 1,2-Dichlorobenzene-d4	11.554	152	59614	5.58	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.60%
Target Compounds						Qvalue
14) Carbon disulfide	2.240	76	8276	0.20	ug/L	95
33) Benzene	5.011	78	48684	1.03	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV021025\
Data File : VV038665.D
Acq On : 10 Feb 2025 14:00
Operator : SY/MD
Sample : Q1335-02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E29C4

Quant Time: Feb 14 11:39:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR020625WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Feb 14 11:36:33 2025
Response via : Initial Calibration

