

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV021025\
 Data File : VV038664.D
 Acq On : 10 Feb 2025 13:36
 Operator : SY/MD
 Sample : Q1335-01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E29C3

Quant Time: Feb 14 11:39:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR020625WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Feb 14 11:36:33 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.526	114	313678	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	306975	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	130384	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	96588	2.98	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	59.60%
7) Chloroethane-d5	1.532	69	77398	3.17	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	63.40%#
11) 1,1-Dichloroethene-d2	2.057	65	41717	3.08	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.60%
20) 2-Butanone-d5	3.799	46	165395	29.36	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	58.72%
24) Chloroform-d	4.240	84	172201	3.35	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	67.00%#
26) 1,2-Dichloroethane-d4	4.934	65	78417	3.46	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	69.20%#
32) Benzene-d6	4.950	84	336676	3.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	75.00%
36) 1,2-Dichloropropane-d6	5.979	67	103144	3.88	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	77.60%
41) Toluene-d8	7.236	98	280530	3.50	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	70.00%
43) trans-1,3-Dichloroprop...	7.555	79	30323	3.10	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	62.00%
46) 2-Hexanone-d5	8.021	63	135318	35.76	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	71.52%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	72821	3.22	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	64.40%#
66) 1,2-Dichlorobenzene-d4	11.555	152	97235	4.30	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	86.00%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV021025\
Data File : VV038664.D
Acq On : 10 Feb 2025 13:36
Operator : SY/MD
Sample : Q1335-01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E29C3

Quant Time: Feb 14 11:39:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR020625WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Feb 14 11:36:33 2025
Response via : Initial Calibration

