

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV021025\
 Data File : VV038668.D
 Acq On : 10 Feb 2025 15:15
 Operator : SY/MD
 Sample : Q1335-05
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E29C8

Quant Time: Feb 14 11:40:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR020625WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Feb 14 11:36:33 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	231217	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	237695	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	98573	5.00	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	89858	3.76	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.20%
7) Chloroethane-d5	1.532	69	67952	3.77	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	75.40%
11) 1,1-Dichloroethene-d2	2.056	65	36115	3.61	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.20%
20) 2-Butanone-d5	3.796	46	162516	39.14	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	78.28%
24) Chloroform-d	4.243	84	149362	3.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.80%
26) 1,2-Dichloroethane-d4	4.937	65	73351	4.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.00%
32) Benzene-d6	4.953	84	299772	4.31	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.20%
36) 1,2-Dichloropropane-d6	5.982	67	91652	4.45	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.00%
41) Toluene-d8	7.236	98	242692	3.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.20%
43) trans-1,3-Dichloroprop...	7.561	79	23413	3.10	ug/L	0.01
Spiked Amount	5.000	Range	55 - 130	Recovery	=	62.00%
46) 2-Hexanone-d5	8.021	63	134842	46.02	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.04%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	71170	4.06	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	81.20%
66) 1,2-Dichlorobenzene-d4	11.554	152	85554	5.01	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.20%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV021025\
Data File : VV038668.D
Acq On : 10 Feb 2025 15:15
Operator : SY/MD
Sample : Q1335-05
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E29C8

Quant Time: Feb 14 11:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR020625WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Feb 14 11:36:33 2025
Response via : Initial Calibration

