

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV021120\
 Data File : VV014556.D
 Acq On : 11 Feb 2020 16:05
 Operator : SY/MD
 Sample : VV0211WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK69

Quant Time: Feb 12 03:25:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM020120WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Feb 12 03:16:31 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	304712	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	287137	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	141193	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	62009	36.64	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	73.28%
7) Chloroethane-d5	1.59	69	46384	44.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	88.16%
11) 1,1-Dichloroethene-d2	2.13	63	79408	35.15	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.30%
21) 2-Butanone-d5	3.92	46	117877	103.52	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.52%
24) Chloroform-d	4.40	84	185096	51.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.36%
26) 1,2-Dichloroethane-d4	5.08	65	121735	56.43	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	112.86%
32) Benzene-d6	5.10	84	389062	51.46	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.92%
36) 1,2-Dichloropropane-d6	6.11	67	121356	50.73	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.46%
41) Toluene-d8	7.36	98	365221	52.33	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.66%
43) trans-1,3-Dichloropropene-	7.66	79	62840	51.69	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	103.38%
47) 2-Hexanone-d5	8.13	63	99874	106.72	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	106.72%
57) 1,1,2,2-Tetrachloroethane-	10.25	84	165553	53.72	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	107.44%
64) 1,2-Dichlorobenzene-d4	11.67	152	143008	54.99	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.98%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV021120\
Data File : VV014556.D
Acq On : 11 Feb 2020 16:05
Operator : SY/MD
Sample : VV0211WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK69

Quant Time: Feb 12 03:25:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM020120WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Feb 12 03:16:31 2020
Response via : Initial Calibration

