

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV021220\
 Data File : VV014572.D
 Acq On : 12 Feb 2020 15:17
 Operator : SY/MD
 Sample : L1489-07
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFRE3

Quant Time: Feb 13 06:28:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM021220WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Feb 13 05:55:11 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	326690	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	307403	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	151925	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	57794	46.07	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	92.14%
7) Chloroethane-d5	1.59	69	42522	48.19	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	96.38%
11) 1,1-Dichloroethene-d2	2.13	63	74659	35.66	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.32%
21) 2-Butanone-d5	3.92	46	96785	93.03	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	93.03%
24) Chloroform-d	4.40	84	172529	47.15	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.30%
26) 1,2-Dichloroethane-d4	5.08	65	110376	48.63	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.26%
32) Benzene-d6	5.10	84	360384	49.02	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.04%
36) 1,2-Dichloropropane-d6	6.11	67	107857	48.71	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.42%
41) Toluene-d8	7.36	98	336132	47.53	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.06%
43) trans-1,3-Dichloropropene-	7.66	79	57890	47.41	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.82%
47) 2-Hexanone-d5	8.13	63	76492	86.45	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	86.45%
57) 1,1,2,2-Tetrachloroethane-	10.25	84	143793	47.85	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.70%
64) 1,2-Dichlorobenzene-d4	11.67	152	130843	48.75	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.50%

Target Compounds

					Ovalue
13) Acetone	2.19	43	4449	4.985 ug/L	98
34) Trichloroethene	5.96	95	4441	2.039 ug/L	79
53) m,p-Xylene	9.18	106	7575	1.943 ug/L	98
54) o-xylene	9.58	106	2968	0.797 ug/L	97
61) Bromoform	9.77	173	1419	0.863 ug/L #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV021220\
Data File : VV014572.D
Acq On : 12 Feb 2020 15:17
Operator : SY/MD
Sample : L1489-07
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFRE3

Quant Time: Feb 13 06:28:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM021220WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Feb 13 05:55:11 2020
Response via : Initial Calibration

