




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021224\
Data File : VV034137.D
Acq On : 12 Feb 2024 14:09
Operator : SY/MD
Sample : P1455-04
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AD7

Manual IntegrationsAPPROVED

Quant Time: Feb 13 00:47:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Feb 13 00:43:48 2024
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 02/14/2024
Supervised By :Semsettin Yesilyurt 02/14/2024



(22) cis-1,2-Dichloroethene (T)

3.815min (+ 0.000) 0.21 ug/L

response 1212

Ion	Exp%	Act%
96.00	100.00	100.00
61.05	129.00	111.65
98.00	61.70	61.20
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021224\
Data File : VV034137.D
Acq On : 12 Feb 2024 14:09
Operator : SY/MD
Sample : P1455-04
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AD7

Manual IntegrationsAPPROVED

Quant Time: Feb 13 00:47:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Feb 13 00:43:48 2024
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 02/14/2024
Supervised By :Semsettin Yesilyurt 02/14/2024



(22) cis-1,2-Dichloroethene (T)

3.828min (+ 0.013) 0.46 ug/L m

response 2693

Ion	Exp%	Act%
96.00	100.00	100.00
61.05	129.00	112.61
98.00	61.70	58.66
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021224\
 Data File : VV034137.D
 Acq On : 12 Feb 2024 14:09
 Operator : SY/MD
 Sample : P1455-04
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AD7

Manual IntegrationsAPPROVED

Quant Time: Feb 13 00:47:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Feb 13 00:43:48 2024
 Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 02/14/2024
 Supervised By :Semsettin Yesilyurt 02/14/2024

Compound	R.T.	Q	Ion	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene	5.539	114		72830	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117		69058	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152		36139	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3	1.278	65		23947	4.246	ug/L	0.00
Spiked Amount 5.000	Range 40	- 130		Recovery	=	85.000%	
7) Chloroethane-d5	1.536	69		20952	4.241	ug/L	0.00
Spiked Amount 5.000	Range 65	- 130		Recovery	=	84.800%	
11) 1,1-Dichloroethene-d2	2.060	65		11563	4.251	ug/L	0.00
Spiked Amount 5.000	Range 60	- 125		Recovery	=	85.000%	
20) 2-Butanone-d5	3.802	46		40364	42.609	ug/L	0.00
Spiked Amount 50.000	Range 40	- 130		Recovery	=	85.220%	
24) Chloroform-d	4.252	84		45803	4.469	ug/L	0.00
Spiked Amount 5.000	Range 70	- 125		Recovery	=	89.400%	
26) 1,2-Dichloroethane-d4	4.947	65		23619	4.931	ug/L	0.00
Spiked Amount 5.000	Range 70	- 130		Recovery	=	98.600%	
32) Benzene-d6	4.963	84		86975	4.198	ug/L	0.00
Spiked Amount 5.000	Range 70	- 125		Recovery	=	84.000%	
36) 1,2-Dichloropropane-d6	5.992	67		24998	4.316	ug/L	0.00
Spiked Amount 5.000	Range 60	- 140		Recovery	=	86.400%	
41) Toluene-d8	7.246	98		82583	4.337	ug/L	0.00
Spiked Amount 5.000	Range 70	- 130		Recovery	=	86.800%	
43) trans-1,3-Dichloroprop...	7.561	79		9513	4.350	ug/L	0.00
Spiked Amount 5.000	Range 55	- 130		Recovery	=	87.000%	
46) 2-Hexanone-d5	8.027	63		33931	43.463	ug/L	0.00
Spiked Amount 50.000	Range 45	- 130		Recovery	=	86.920%	
56) 1,1,2,2-Tetrachloroeth...	10.153	84		16168	4.605	ug/L	0.00
Spiked Amount 5.000	Range 65	- 120		Recovery	=	92.000%	
66) 1,2-Dichlorobenzene-d4	11.561	152		30012	4.566	ug/L	0.00
Spiked Amount 5.000	Range 80	- 120		Recovery	=	91.400%	
Target Compounds							
12) 1,1-Dichloroethene	2.069	96		1324	0.260	ug/L	# 1
22) cis-1,2-Dichloroethene	3.828	96		2693m	0.458	ug/L	
25) Chloroform	4.278	83		7105	0.738	ug/L	100
31) Carbon tetrachloride	4.738	117		23272	2.869	ug/L	98
34) Trichloroethene	5.838	95		10230	1.707	ug/L	96
47) Tetrachloroethene	7.905	164		14241	2.924	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021224\
Data File : VV034137.D
Acq On : 12 Feb 2024 14:09
Operator : SY/MD
Sample : P1455-04
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AD7

Manual IntegrationsAPPROVED

Quant Time: Feb 13 00:47:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Feb 13 00:43:48 2024
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 02/14/2024
Supervised By :Semsettin Yesilyurt 02/14/2024

