

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW021224\
 Data File : VW034133.D
 Acq On : 12 Feb 2024 12:35
 Operator : SY/MD
 Sample : VIBLK320
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK320

Quant Time: Feb 13 00:46:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Feb 13 00:43:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	80565	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	74906	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	38562	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	25440	4.077	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.600%
7) Chloroethane-d5	1.535	69	22205	4.063	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	81.200%
11) 1,1-Dichloroethene-d2	2.060	65	12213	4.059	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	81.200%
20) 2-Butanone-d5	3.818	46	40027	38.197	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	76.400%
24) Chloroform-d	4.252	84	47527	4.192	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.800%
26) 1,2-Dichloroethane-d4	4.947	65	25216	4.759	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.200%
32) Benzene-d6	4.963	84	93840	4.176	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.600%
36) 1,2-Dichloropropane-d6	5.992	67	25605	4.076	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	81.600%
41) Toluene-d8	7.246	98	87310	4.227	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.600%
43) trans-1,3-Dichloroprop...	7.561	79	9193	3.875	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	77.600%
46) 2-Hexanone-d5	8.027	63	36002	42.516	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.040%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	16380	4.301	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.000%
66) 1,2-Dichlorobenzene-d4	11.561	152	30519	4.352	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	87.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021224\
Data File : VV034133.D
Acq On : 12 Feb 2024 12:35
Operator : SY/MD
Sample : VIBLK320
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK320

Quant Time: Feb 13 00:46:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Feb 13 00:43:48 2024
Response via : Initial Calibration

