

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW021521\
 Data File : VW020324.D
 Acq On : 15 Feb 2021 13:27
 Operator : SY/MD
 Sample : VIBLK91
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK91

Quant Time: Feb 16 00:34:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM020521WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Feb 16 00:32:19 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	727491	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	694549	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.255	152	349519	50.00	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	161346	30.42	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	60.84%
7) Chloroethane-d5	1.568	69	149199	34.29	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	68.58%#
11) 1,1-Dichloroethene-d2	2.108	63	254697	25.40	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	50.80%#
21) 2-Butanone-d5	3.879	46	241678	81.56	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	81.56%
24) Chloroform-d	4.352	84	330951	35.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	70.12%
26) 1,2-Dichloroethane-d4	5.037	65	231834	38.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	77.48%
32) Benzene-d6	5.053	84	674896	37.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	75.00%
36) 1,2-Dichloropropane-d6	6.072	67	210888	39.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	78.52%
41) Toluene-d8	7.320	98	612978	36.25	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	72.50%#
43) trans-1,3-Dichloroprop...	7.625	79	91807	33.53	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.06%
47) 2-Hexanone-d5	8.091	63	160227	71.69	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	71.69%
57) 1,1,2,2-Tetrachloroeth...	10.220	84	281491	36.49	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	72.98%
64) 1,2-Dichlorobenzene-d4	11.631	152	264930	40.06	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	80.12%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021521\
Data File : VV020324.D
Acq On : 15 Feb 2021 13:27
Operator : SY/MD
Sample : VIBLK91
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK91

Quant Time: Feb 16 00:34:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM020521WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Feb 16 00:32:19 2021
Response via : Initial Calibration

