

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021523\
 Data File : VV030050.D
 Acq On : 15 Feb 2023 13:23
 Operator : SY/MD
 Sample : VSTD00224
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD002224

Quant Time: Feb 16 01:32:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM021523.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Feb 16 01:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	20381	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	18885	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	10737	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	32866	1.745	ug/L	0.00
4) 1,2-Dichloroethane-d4	4.946	65	23419	2.028	ug/L	0.00
7) 1,2-Dichloropropane-d6	5.995	67	29861	1.994	ug/L	0.00
8) Toluene-d8	7.238	98	88058	2.053	ug/L	-0.02
10) 1,1,2,2-Tetrachloroeth...	10.146	84	21141	2.128	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.280	62	30634	1.715	ug/L	92
6) Trichloroethene	5.828	95	27283	1.927	ug/L	86
9) 1,2-Dibromoethane	8.290	107	18400	2.205	ug/L #	99
12) 1,2,3-Trichloropropane	10.214	75	15657	1.901	ug/L	96
13) 1,2-Dibromo-3-chloropr...	12.357	75	3520	2.307	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021523\
Data File : VV030050.D
Acq On : 15 Feb 2023 13:23
Operator : SY/MD
Sample : VSTD00224
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD002224

Quant Time: Feb 16 01:32:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM021523.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Feb 16 01:30:38 2023
Response via : Initial Calibration

