

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021524\
 Data File : VV034185.D
 Acq On : 15 Feb 2024 13:11
 Operator : SY/MD
 Sample : VV0215WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK216

Quant Time: Feb 16 00:41:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Feb 16 00:40:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	77307	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	71889	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	36429	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	21552	3.600	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	72.000%
7) Chloroethane-d5	1.535	69	20778	3.962	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	79.200%
11) 1,1-Dichloroethene-d2	2.060	65	11158	3.864	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.200%
20) 2-Butanone-d5	3.831	46	40759	40.534	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	81.060%
24) Chloroform-d	4.249	84	47397	4.357	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.200%
26) 1,2-Dichloroethane-d4	4.947	65	24301	4.780	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.600%
32) Benzene-d6	4.960	84	88107	4.085	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.800%
36) 1,2-Dichloropropane-d6	5.989	67	26329	4.367	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.400%
41) Toluene-d8	7.243	98	80767	4.074	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.400%
43) trans-1,3-Dichloroprop...	7.561	79	9296	4.083	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.600%
46) 2-Hexanone-d5	8.030	63	33392	41.088	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.180%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	16403	4.488	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.800%
66) 1,2-Dichlorobenzene-d4	11.561	152	30431	4.593	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021524\
Data File : VV034185.D
Acq On : 15 Feb 2024 13:11
Operator : SY/MD
Sample : VV0215WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK216

Quant Time: Feb 16 00:41:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Feb 16 00:40:49 2024
Response via : Initial Calibration

