

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021524\
 Data File : VV034194.D
 Acq On : 15 Feb 2024 17:55
 Operator : SY/MD
 Sample : P1501-01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E1098

Quant Time: Feb 16 00:43:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Feb 16 00:40:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	71045	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	65806	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	31751	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	19171	3.484	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	69.600%	
7) Chloroethane-d5	1.536	69	19094	3.962	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	79.200%	
11) 1,1-Dichloroethene-d2	2.060	65	10122	3.815	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	76.200%	
20) 2-Butanone-d5	3.806	46	48116	52.069	ug/L	-0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	=	104.140%	
24) Chloroform-d	4.253	84	47134	4.714	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	94.200%	
26) 1,2-Dichloroethane-d4	4.947	65	25237	5.401	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	108.000%	
32) Benzene-d6	4.963	84	82527	4.180	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	83.600%	
36) 1,2-Dichloropropane-d6	5.989	67	23942	4.338	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	86.800%	
41) Toluene-d8	7.246	98	74739	4.119	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	82.400%	
43) trans-1,3-Dichloroprop...	7.558	79	9408	4.515	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	90.200%	
46) 2-Hexanone-d5	8.027	63	37572	50.505	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	101.020%	
56) 1,1,2,2-Tetrachloroeth...	10.153	84	16781	5.016	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	100.400%	
66) 1,2-Dichlorobenzene-d4	11.561	152	27389	4.743	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	94.800%	
Target Compounds						Qvalue
34) Trichloroethene	5.838	95	5709	0.999	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021524\
Data File : VV034194.D
Acq On : 15 Feb 2024 17:55
Operator : SY/MD
Sample : P1501-01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E1098

Quant Time: Feb 16 00:43:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Feb 16 00:40:49 2024
Response via : Initial Calibration

