

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW021524\
 Data File : VW034187.D
 Acq On : 15 Feb 2024 14:18
 Operator : SY/MD
 Sample : P1498-09
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AH2

Quant Time: Feb 16 00:42:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Feb 16 00:40:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	80245	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	76353	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	38863	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	19515	3.140	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	62.800%
7) Chloroethane-d5	1.536	69	19350	3.555	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	71.000%
11) 1,1-Dichloroethene-d2	2.060	65	10107	3.372	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.400%
20) 2-Butanone-d5	3.831	46	41040	39.320	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	78.640%
24) Chloroform-d	4.252	84	46257	4.096	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.000%
26) 1,2-Dichloroethane-d4	4.947	65	23618	4.475	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.600%
32) Benzene-d6	4.963	84	82769	3.613	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	72.200%
36) 1,2-Dichloropropane-d6	5.992	67	24381	3.807	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	76.200%
41) Toluene-d8	7.243	98	76235	3.621	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	72.400%
43) trans-1,3-Dichloroprop...	7.561	79	8924	3.691	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.800%
46) 2-Hexanone-d5	8.027	63	31234	36.186	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	72.380%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	15343	3.952	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	79.000%
66) 1,2-Dichlorobenzene-d4	11.561	152	28336	4.009	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	80.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021524\
Data File : VV034187.D
Acq On : 15 Feb 2024 14:18
Operator : SY/MD
Sample : P1498-09
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COAH2

Quant Time: Feb 16 00:42:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Feb 16 00:40:49 2024
Response via : Initial Calibration

