

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV021720\
 Data File : VV014595.D
 Acq On : 17 Feb 2020 15:59
 Operator : SY/MD
 Sample : VV0217MBL07
 Misc : 5.0µ/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK214

Quant Time: Feb 18 06:59:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVLM021720W.M
 Quant Title : VOC Analysis
 QLast Update : Tue Feb 18 06:41:33 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	294347	50.00	µg/L	0.00
28) Chlorobenzene-d5	8.89	117	274493	50.00	µg/L	0.00
58) 1,4-Dichlorobenzene-d4	11.29	152	133511	50.00	µg/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	44929	45.84	µg/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	91.68%
7) Chloroethane-d5	1.58	69	35022	47.15	µg/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.30%
11) 1,1-Dichloroethene-d2	2.13	63	63795	35.24	µg/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.48%
21) 2-Butanone-d5	3.92	46	90924	95.31	µg/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	95.31%
24) Chloroform-d	4.40	84	152211	47.08	µg/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.16%
26) 1,2-Dichloroethane-d4	5.08	65	100476	49.05	µg/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.10%
32) Benzene-d6	5.10	84	315096	50.16	µg/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.32%
36) 1,2-Dichloropropane-d6	6.11	67	96742	50.95	µg/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.90%
41) Toluene-d8	7.35	98	300674	50.29	µg/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.58%
43) trans-1,3-Dichloropropene-	7.66	79	52273	49.02	µg/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.04%
47) 2-Hexanone-d5	8.13	63	73328	97.79	µg/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.79%
56) 1,1,2,2-Tetrachloroethane-	10.25	84	130361	49.23	µg/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.46%
66) 1,2-Dichlorobenzene-d4	11.67	152	119112	52.44	µg/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.88%

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV021720\
 Data File : VV014595.D
 Acq On : 17 Feb 2020 15:59
 Operator : SY/MD
 Sample : VV0217MBL07
 Misc : 5.0µ/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampled :
 VBLK214

Quant Time: Feb 18 06:59:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVLM021720W.M
 Quant Title : VOC Analysis
 QLast Update : Tue Feb 18 06:41:33 2020
 Response via : Initial Calibration

