

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV021920\
 Data File : VV014612.D
 Acq On : 19 Feb 2020 17:47
 Operator : SY/MD
 Sample : VV0219WBL01
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK701

Quant Time: Feb 20 02:57:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVTR021920WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Feb 20 02:48:04 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	249722	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	232418	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.29	152	112477	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	43283	4.72	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.40%
7) Chloroethane-d5	1.58	69	34148	4.84	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.80%
11) 1,1-Dichloroethene-d2	2.13	63	61981	3.59	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.80%
20) 2-Butanone-d5	3.95	46	88263	45.72	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	91.44%
24) Chloroform-d	4.40	84	90185	4.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.60%
26) 1,2-Dichloroethane-d4	5.08	65	70176	5.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.60%
32) Benzene-d6	5.09	84	292739	4.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.20%
36) 1,2-Dichloropropane-d6	6.11	67	83619	4.97	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.40%
41) Toluene-d8	7.35	98	275320	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.00%
43) trans-1,3-Dichloropropene-	7.66	79	33965	4.77	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.40%
46) 2-Hexanone-d5	8.14	63	125922	51.64	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.28%
56) 1,1,2,2-Tetrachloroethane-	10.25	84	59813	4.79	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.80%
66) 1,2-Dichlorobenzene-d4	11.67	152	97632	4.91	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.20%

Target Compounds					Ovalue
25) Chloroform	4.42	83	53346	1.353 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV021920\
Data File : VV014612.D
Acq On : 19 Feb 2020 17:47
Operator : SY/MD
Sample : VV0219WBL01
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK701

Quant Time: Feb 20 02:57:04 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVTR021920WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Feb 20 02:48:04 2020
Response via : Initial Calibration

