

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV021920\
 Data File : VV014613.D
 Acq On : 19 Feb 2020 18:11
 Operator : SY/MD
 Sample : VV0219WBL02
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK702

Quant Time: Feb 20 03:00:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVTR021920WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Feb 20 02:48:04 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	243631	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	227231	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.29	152	110571	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	44334	4.95	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	99.00%
7) Chloroethane-d5	1.58	69	34843	5.07	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.40%
11) 1,1-Dichloroethene-d2	2.12	63	63894	3.80	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.00%
20) 2-Butanone-d5	3.95	46	99193	52.67	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.34%
24) Chloroform-d	4.40	84	93653	4.29	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.80%
26) 1,2-Dichloroethane-d4	5.07	65	71908	5.28	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.60%
32) Benzene-d6	5.09	84	305836	5.25	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.00%
36) 1,2-Dichloropropane-d6	6.11	67	86917	5.28	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.60%
41) Toluene-d8	7.35	98	285728	5.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.80%
43) trans-1,3-Dichloropropene-	7.66	79	35462	5.09	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	101.80%
46) 2-Hexanone-d5	8.13	63	130207	54.61	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	109.22%
56) 1,1,2,2-Tetrachloroethane-	10.25	84	63080	5.17	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.40%
66) 1,2-Dichlorobenzene-d4	11.67	152	101538	5.19	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.80%

Target Compounds				Ovalue
25) Chloroform	4.42	83	56762	1.476 ug/L 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV021920\
Data File : VV014613.D
Acq On : 19 Feb 2020 18:11
Operator : SY/MD
Sample : VV0219WBL02
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK702

Quant Time: Feb 20 03:00:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVTR021920WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Feb 20 02:48:04 2020
Response via : Initial Calibration

