

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW021924\
 Data File : VW034249.D
 Acq On : 19 Feb 2024 17:32
 Operator : SY/MD
 Sample : P1503-16
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DCNC3

Quant Time: Feb 20 03:30:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Feb 20 01:47:35 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	69692	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	64492	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	32598	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	22133	4.101	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.000%
7) Chloroethane-d5	1.532	69	19764	4.181	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	83.600%
11) 1,1-Dichloroethene-d2	2.056	65	11292	4.338	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	86.800%
20) 2-Butanone-d5	3.834	46	29660	32.720	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	65.440%
24) Chloroform-d	4.252	84	43275	4.412	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.200%
26) 1,2-Dichloroethane-d4	4.947	65	22399	4.887	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.800%
32) Benzene-d6	4.963	84	80998	4.186	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.800%
36) 1,2-Dichloropropane-d6	5.992	67	22078	4.082	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	81.600%
41) Toluene-d8	7.246	98	76587	4.307	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.200%
43) trans-1,3-Dichloroprop...	7.561	79	8229	4.029	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.600%
46) 2-Hexanone-d5	8.030	63	34037	46.686	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.380%
56) 1,1,2,2-Tetrachloroeth...	10.152	84	15402	4.697	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.000%
66) 1,2-Dichlorobenzene-d4	11.561	152	28269	4.768	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.400%
Target Compounds						Qvalue
17) Methyl tert-butyl Ether	2.706	73	167317	16.726	ug/L	97
22) cis-1,2-Dichloroethene	3.828	96	1239	0.220	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021924\
Data File : VV034249.D
Acq On : 19 Feb 2024 17:32
Operator : SY/MD
Sample : P1503-16
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DCNC3

Quant Time: Feb 20 03:30:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Feb 20 01:47:35 2024
Response via : Initial Calibration

