

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW021924\
 Data File : VW034233.D
 Acq On : 19 Feb 2024 11:01
 Operator : SY/MD
 Sample : P1513-03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0U22

Quant Time: Feb 19 11:46:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Feb 19 00:41:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	73756	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	67758	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	33547	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	23088	4.042	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.800%
7) Chloroethane-d5	1.535	69	20618	4.121	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.400%
11) 1,1-Dichloroethene-d2	2.059	65	11644	4.227	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	84.600%
20) 2-Butanone-d5	3.841	46	32184	33.548	ug/L	0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	67.100%
24) Chloroform-d	4.252	84	44248	4.263	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.200%
26) 1,2-Dichloroethane-d4	4.947	65	22971	4.736	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.800%
32) Benzene-d6	4.960	84	84308	4.147	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.000%
36) 1,2-Dichloropropane-d6	5.989	67	23221	4.086	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	81.800%
41) Toluene-d8	7.243	98	77941	4.172	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.400%
43) trans-1,3-Dichloroprop...	7.561	79	8828	4.114	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.200%
46) 2-Hexanone-d5	8.030	63	27050	35.314	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	70.620%
56) 1,1,2,2-Tetrachloroeth...	10.152	84	14247	4.136	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.800%
66) 1,2-Dichlorobenzene-d4	11.561	152	26865	4.403	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	88.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV021924\
Data File : VV034233.D
Acq On : 19 Feb 2024 11:01
Operator : SY/MD
Sample : P1513-03
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0U22

Quant Time: Feb 19 11:46:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Feb 19 00:41:47 2024
Response via : Initial Calibration

