

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW022124\
 Data File : VW034316.D
 Acq On : 21 Feb 2024 13:54
 Operator : SY/MD
 Sample : P1503-09
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DCNB6

Quant Time: Feb 22 00:21:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Feb 22 00:18:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	70019	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	65037	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	32561	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	19030	3.509	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	70.200%	
7) Chloroethane-d5	1.532	69	18057	3.802	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	76.000%	
11) 1,1-Dichloroethene-d2	2.060	65	9756	3.731	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	74.600%	
20) 2-Butanone-d5	3.799	46	47062	51.674	ug/L	-0.04
Spiked Amount	50.000	Range 40 - 130	Recovery	=	103.340%	
24) Chloroform-d	4.252	84	43394	4.404	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	88.000%	
26) 1,2-Dichloroethane-d4	4.947	65	23194	5.037	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	100.800%	
32) Benzene-d6	4.960	84	79322	4.065	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	81.400%	
36) 1,2-Dichloropropane-d6	5.989	67	24207	4.438	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	88.800%	
41) Toluene-d8	7.243	98	71862	4.007	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	80.200%	
43) trans-1,3-Dichloroprop...	7.561	79	6714	3.260	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	65.200%	
46) 2-Hexanone-d5	8.027	63	35226	47.912	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	95.820%	
56) 1,1,2,2-Tetrachloroeth...	10.153	84	17067	5.161	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	103.200%	
66) 1,2-Dichlorobenzene-d4	11.561	152	25939	4.380	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	87.600%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022124\
Data File : VV034316.D
Acq On : 21 Feb 2024 13:54
Operator : SY/MD
Sample : P1503-09
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DCNB6

Quant Time: Feb 22 00:21:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Feb 22 00:18:57 2024
Response via : Initial Calibration

