

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022124\
 Data File : VV034313.D
 Acq On : 21 Feb 2024 12:40
 Operator : SY/MD
 Sample : P1513-07
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK002

Quant Time: Feb 22 00:21:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Feb 22 00:18:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	80207	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	73513	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	35210	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	21480	3.458	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	69.200%	
7) Chloroethane-d5	1.536	69	21125	3.883	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	77.600%	
11) 1,1-Dichloroethene-d2	2.060	65	11090	3.702	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	74.000%	
20) 2-Butanone-d5	3.831	46	41757	40.025	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	80.060%	
24) Chloroform-d	4.253	84	50401	4.465	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	89.400%	
26) 1,2-Dichloroethane-d4	4.947	65	25524	4.839	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	96.800%	
32) Benzene-d6	4.960	84	92662	4.201	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	84.000%	
36) 1,2-Dichloropropane-d6	5.989	67	26450	4.290	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	85.800%	
41) Toluene-d8	7.243	98	83348	4.112	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	82.200%	
43) trans-1,3-Dichloroprop...	7.564	79	7411	3.183	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	63.600%	
46) 2-Hexanone-d5	8.031	63	35224	42.385	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	84.780%	
56) 1,1,2,2-Tetrachloroeth...	10.153	84	16913	4.525	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	90.600%	
66) 1,2-Dichlorobenzene-d4	11.561	152	29098	4.544	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	90.800%	
Target Compounds						Qvalue
16) Methylene chloride	2.452	84	1304	0.243	ug/L	92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022124\
Data File : VV034313.D
Acq On : 21 Feb 2024 12:40
Operator : SY/MD
Sample : P1513-07
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK002

Quant Time: Feb 22 00:21:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Feb 22 00:18:57 2024
Response via : Initial Calibration





#16
Methylene chloride
Concen: 0.243 ug/L
RT: 2.452 min Scan# 41
Delta R.T. 0.003 min
Lab File: VV034313.D
Acq: 21 Feb 2024 12:40

Instrument :
MSVOA_V
ClientSampleId :
VHBLK002

Tgt Ion:	84	Resp:	1304
Ion	Ratio	Lower	Upper
84	100		
86	49.4	44.7	83.1
49	129.9	93.7	173.9

