

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW022324\
 Data File : VW034363.D
 Acq On : 23 Feb 2024 10:06
 Operator : SY/MD
 Sample : P1498-18DL 10X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AC2DL

Quant Time: Feb 24 00:45:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Feb 24 00:41:32 2024
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		5.535	114	78368	5.000	ug/L	0.00
28) Chlorobenzene-d5		8.786	117	71261	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.188	152	34519	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.278	65	31784	5.237	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	104.800%	
7) Chloroethane-d5		1.532	69	26094	4.908	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.200%	
11) 1,1-Dichloroethene-d2		2.060	65	14637	5.001	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	100.000%	
20) 2-Butanone-d5		3.822	46	32782	32.160	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	64.320%	
24) Chloroform-d		4.252	84	55449	5.028	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.600%	
26) 1,2-Dichloroethane-d4		4.947	65	25957	5.036	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.800%	
32) Benzene-d6		4.960	84	105901	4.953	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.000%	
36) 1,2-Dichloropropane-d6		5.989	67	29397	4.919	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.400%	
41) Toluene-d8		7.243	98	95850	4.878	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.600%	
43) trans-1,3-Dichloroprop...		7.561	79	9142	4.051	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.000%	
46) 2-Hexanone-d5		8.030	63	33978	42.178	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.360%	
56) 1,1,2,2-Tetrachloroeth...		10.156	84	17093	4.718	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.400%	
66) 1,2-Dichlorobenzene-d4		11.561	152	30711	4.892	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.800%	
Target Compounds							
						Qvalue	
12) 1,1-Dichloroethene		2.069	96	4281	0.782	ug/L	# 1
16) Methylene chloride		2.452	84	1156	0.220	ug/L	77
22) cis-1,2-Dichloroethene		3.825	96	2375	0.375	ug/L	92
34) Trichloroethene		5.834	95	53782	8.695	ug/L	99
47) Tetrachloroethene		7.908	164	1127	0.224	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022324\
Data File : VV034363.D
Acq On : 23 Feb 2024 10:06
Operator : SY/MD
Sample : P1498-18DL 10X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AC2DL

Quant Time: Feb 24 00:45:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Feb 24 00:41:32 2024
Response via : Initial Calibration

