

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022324\
 Data File : VV034364.D
 Acq On : 23 Feb 2024 10:30
 Operator : SY/MD
 Sample : P1498-10
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Feb 24 00:45:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Feb 24 00:41:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	81985	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	77564	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	36633	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	32985	5.195	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	103.800%	
7) Chloroethane-d5	1.536	69	27633	4.969	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	99.400%	
11) 1,1-Dichloroethene-d2	2.060	65	15741	5.141	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	102.800%	
20) 2-Butanone-d5	3.822	46	35147	32.959	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	65.920%	
24) Chloroform-d	4.253	84	57409	4.976	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	99.600%	
26) 1,2-Dichloroethane-d4	4.947	65	28016	5.196	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	104.000%	
32) Benzene-d6	4.960	84	111466	4.790	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	95.800%	
36) 1,2-Dichloropropane-d6	5.992	67	31335	4.817	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	96.400%	
41) Toluene-d8	7.246	98	104446	4.884	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	97.600%	
43) trans-1,3-Dichloroprop...	7.561	79	9375	3.817	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	76.400%	
46) 2-Hexanone-d5	8.031	63	38208	43.575	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	87.140%	
56) 1,1,2,2-Tetrachloroeth...	10.156	84	19610	4.973	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	99.400%	
66) 1,2-Dichlorobenzene-d4	11.561	152	34232	5.138	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	102.800%	
Target Compounds						Qvalue
16) Methylene chloride	2.449	84	918	0.167	ug/L	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022324\
Data File : VV034364.D
Acq On : 23 Feb 2024 10:30
Operator : SY/MD
Sample : P1498-10
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Feb 24 00:45:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Feb 24 00:41:32 2024
Response via : Initial Calibration

