

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW022324\
 Data File : VW034367.D
 Acq On : 23 Feb 2024 11:44
 Operator : SY/MD
 Sample : VIBLK326
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK326

Quant Time: Feb 24 00:46:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Feb 24 00:41:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	27611	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	26402	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	12784	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	15997	7.481	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 149.600%#		
7) Chloroethane-d5	1.526	69	9525	5.085	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 101.800%		
11) 1,1-Dichloroethene-d2	2.060	65	5553	5.385	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 107.600%		
20) 2-Butanone-d5	3.799	46	39029	108.674	ug/L	-0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	= 217.340%#		
24) Chloroform-d	4.256	84	19004	4.891	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 97.800%		
26) 1,2-Dichloroethane-d4	4.950	65	12785	7.041	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 140.800%#		
32) Benzene-d6	4.966	84	35856	4.527	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 90.600%		
36) 1,2-Dichloropropane-d6	5.995	67	10208	4.610	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 92.200%		
41) Toluene-d8	7.249	98	32689	4.490	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 89.800%		
43) trans-1,3-Dichloroprop...	7.567	79	3385	4.049	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 81.000%		
46) 2-Hexanone-d5	8.027	63	30052	100.687	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 201.380%#		
56) 1,1,2,2-Tetrachloroeth...	10.153	84	11293	8.413	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 168.200%#		
66) 1,2-Dichlorobenzene-d4	11.564	152	12136	5.220	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 104.400%		
Target Compounds						Qvalue
3) Chloromethane	1.217	50	2653	0.961	ug/L	98
13) Acetone	2.117	43	2606	9.565	ug/L #	40
14) Carbon disulfide	2.240	76	17749	2.844	ug/L	98
16) Methylene chloride	2.452	84	3240	1.751	ug/L	92
17) Methyl tert-butyl Ether	2.706	73	60183	15.186	ug/L	98
18) trans-1,2-Dichloroethene	2.693	96	86814	44.188	ug/L	96
19) 1,1-Dichloroethane	3.114	63	3270	0.906	ug/L	96
22) cis-1,2-Dichloroethene	3.818	96	15157	6.800	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022324\
Data File : VV034367.D
Acq On : 23 Feb 2024 11:44
Operator : SY/MD
Sample : VIBLK326
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK326

Quant Time: Feb 24 00:46:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Feb 24 00:41:32 2024
Response via : Initial Calibration

