

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022324\
 Data File : VV034362.D
 Acq On : 23 Feb 2024 09:40
 Operator : SY/MD
 Sample : VV0223WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK223

Quant Time: Feb 24 00:45:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Feb 24 00:41:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	82679	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	74908	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	35825	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	30373	4.743	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.800%
7) Chloroethane-d5	1.532	69	25638	4.571	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.400%
11) 1,1-Dichloroethene-d2	2.060	65	14233	4.609	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	92.200%
20) 2-Butanone-d5	3.844	46	40386	37.554	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	75.100%
24) Chloroform-d	4.253	84	51866	4.458	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.200%
26) 1,2-Dichloroethane-d4	4.947	65	26042	4.789	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.800%
32) Benzene-d6	4.960	84	101733	4.527	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.600%
36) 1,2-Dichloropropane-d6	5.989	67	27281	4.343	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.800%
41) Toluene-d8	7.243	98	92665	4.486	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.800%
43) trans-1,3-Dichloroprop...	7.561	79	8221	3.466	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	69.400%
46) 2-Hexanone-d5	8.031	63	34221	40.411	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	80.820%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	17337	4.552	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.000%
66) 1,2-Dichlorobenzene-d4	11.561	152	30378	4.663	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022324\
Data File : VV034362.D
Acq On : 23 Feb 2024 09:40
Operator : SY/MD
Sample : VV0223WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK223

Quant Time: Feb 24 00:45:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Feb 24 00:41:32 2024
Response via : Initial Calibration

