

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW022324\
 Data File : VW034374.D
 Acq On : 23 Feb 2024 14:33
 Operator : SY/MD
 Sample : P1566-15
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 JWAC9

Quant Time: Feb 24 00:48:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Feb 24 00:41:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	78556	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	72681	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	35035	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	33103	5.441	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	108.800%
7) Chloroethane-d5	1.535	69	26400	4.954	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.000%
11) 1,1-Dichloroethene-d2	2.063	65	14657	4.995	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	100.000%
20) 2-Butanone-d5	3.809	46	48561	47.526	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.060%
24) Chloroform-d	4.252	84	54839	4.960	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.200%
26) 1,2-Dichloroethane-d4	4.947	65	27799	5.381	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.600%
32) Benzene-d6	4.963	84	103386	4.741	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.800%
36) 1,2-Dichloropropane-d6	5.992	67	27831	4.566	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.400%
41) Toluene-d8	7.246	98	93556	4.668	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.400%
43) trans-1,3-Dichloroprop...	7.561	79	5972	2.595	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	51.800%#
46) 2-Hexanone-d5	8.027	63	39781	48.417	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.840%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	18369	4.971	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.400%
66) 1,2-Dichlorobenzene-d4	11.561	152	33143	5.202	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022324\
Data File : VV034374.D
Acq On : 23 Feb 2024 14:33
Operator : SY/MD
Sample : P1566-15
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
JWAC9

Quant Time: Feb 24 00:48:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Feb 24 00:41:32 2024
Response via : Initial Calibration

